



Decision Matrix

BATH BED SLIDE TRANSFER BATHING SYSTEM

General:

- Weight of client? Standard unit accommodates 115 lbs and can be customized up to 175 lbs
- Seat Width of client? Standard seat and back width is 18.5".

Backrest: Standard backrest length is 26"

- Do you want the backrest to stop above the clients head for full support? This will accommodate head bolsters.
- Do you want the backrest to stop at the client's lower head or back of neck for support and ease of hair washing?
- Does the client require a Torso Belt?
- Are Drop-Down Armrests necessary for the client? (Adds 5.5" to overall width)
- Are Lateral Supports necessary for further trunk support?
- Is a Head Bolster necessary for stable head positioning?
- Does the client need an 'H' or "X" Harness?
- Would a padded panel seat and back in place of the tension adjustable mesh benefit the client?

Seating: Standard seat depth is 14".

- Is the standard seat depth adequate for the client's leg length?
- Does the client need support for their lower legs other than the pull-out bar? (Calf Extension)
- Are ankle restraints necessary?

Rolling Chassis: Working height is set at 32"

- Is this working height best for the care giver? It can be modified to client specifications.

Tub Frame:

- Has the manufacturer reviewed the completed bathroom diagram for necessary modification?
- Does the client prefer lock out pins verses D Pins for leg adjustment (requires up-charge)

Convenience:

- Would a Shower Wand Holder or Shower Caddie benefit the client?

Please view our website for YouTube operational video